

EARL PET/CT | PET/MR Accreditation 2026

THIS IS TO CERTIFY THAT THE

Lausanne University Hospital
Department of Nuclear Medicine
Scanner: Siemens Biograph Vision 600 Edge 11119

is granted accreditation according to the requirements and specifications indicated in the EARL accreditation user manual for the following oncology isotope imaging and accreditation program:

^{18}F standards 2, Brain PET/CT

PET/CT | PET/MR scanner accreditation ensures similar performance of the systems within a multicentre setting by harmonising reconstructions for quantification of PET/CT | PET/MR scans. Accredited PET/CT | PET/MR centres of excellence can compare, exchange and combine PET/CT | PET/MR findings, including SUVs, since data are collected and processed in a standardised manner. The regular submission and approval of QC data, as required by EARL, is a condition for the maintenance of the accreditation.

Date of Print: Q1-2026



Paola Anna Erba
EANM President 2025-2026



Ivalina Hristova
Chief Scientific Officer

