

EARL PET/CT | PET/MR Accreditation 2026

THIS IS TO CERTIFY THAT THE

Lausanne University Hospital
Department of Nuclear Medicine
Scanner: Siemens Biograph Vision 600 10001

is granted accreditation according to the requirements and specifications indicated in the EARL accreditation user manual for the following oncology isotope imaging and accreditation program:

^{18}F standards 2, Brain PET/CT

PET/CT | PET/MR scanner accreditation ensures similar performance of the systems within a multicentre setting by harmonising reconstructions for quantification of PET/CT | PET/MR scans. Accredited PET/CT | PET/MR centres of excellence can compare, exchange and combine PET/CT | PET/MR findings, including SUVs, since data are collected and processed in a standardised manner. The regular submission and approval of QC data, as required by EARL, is a condition for the maintenance of the accreditation.

Date of Print: Q1-2026



Paola Anna Erba
EANM President 2025-2026



Ivalina Hristova
Chief Scientific Officer

